

BHCG			
	Quality Control		
	Level 1	Level 2	Level 3
1	3.88	23.42	360.99
2	3.95	19.37	375.92
3	3.02	18.67	394.37
4	2.92	18.72	375.97
5	3.24	24.99	386.42
6	3.17	25.69	262.27
7	3.17	23.9	244.29
8	3.12	25.29	409.01
9	3.13	25.1	394.37
10	3.01	24.72	
11	3.23	23.57	
12	3.13	24.26	
13	4.34	23.6	
14	4.16		
15	3.92		
16	3.83		
17	4.11		
18	3.23		
19	3.13		
20	4.34		
21	4.25		
22			
23			
Total	74.28	2805.01	3203.61
Mean	3.54	215.77	355.96
SDI	0.51	695.04	59.96
CV	0.14	3.22	0.17

CLIA acceptance performance for BHCG is 25%.
Acceptable limit performance for RIQAS is 12,8%.