

| CEA | | | |
|--------------|------------------------|----------------|----------------|
| | Quality Control | | |
| | Level 1 | Level 2 | Level 3 |
| 1 | 2 | 19 | 43 |
| 2 | 3 | 19 | 41 |
| 3 | 3 | 19 | 43 |
| 4 | 3 | 18 | 45 |
| 5 | 2 | 18 | 39 |
| 6 | 3 | 20 | 43 |
| 7 | 3 | 20 | 46 |
| 8 | 3 | 20 | 46 |
| 9 | 3 | 18 | 43 |
| 10 | 3 | 18 | 46 |
| 11 | 3 | 17 | 43 |
| 12 | | | 44 |
| 13 | | | 44 |
| 14 | | | 44 |
| 15 | | | 42 |
| 16 | | | 44 |
| 17 | | | |
| 18 | | | |
| 19 | | | |
| 20 | | | |
| 21 | | | |
| 22 | | | |
| 23 | | | |
| Total | 31 | 206 | 696 |
| Mean | 2.82 | 18.73 | 43.50 |
| SDI | 0.40 | 1.01 | 1.86 |
| CV | 0.14 | 0.05 | 0.04 |

CLIA acceptance performance for CEA is 25%.

Acceptable limit performance for RIQAS is 11,3%.