

FSH			
	Quality Control		FSH
	Level 1	Level 2	Level 3
1	9.9	20.29	36.23
2	8.58	17.5	32.72
3	9.85	19.45	34.72
4	9.07	18.54	36.06
5	9.7	20.69	34.63
6	9.81	20.17	36.8
7	9.43	19.71	35.27
8	8.99	18.25	36.56
9	9.25	18.81	35.42
10	9.19	19.5	34.94
11	9.31	19.49	34.2
12	8.62	19.27	34.4
13	8.76	18.72	33.2
14	8.4	18.73	33.53
15	9.09	18.66	33.98
16	8.69	17.36	33.57
17	8.95	19.04	33.69
18	9.76	19.55	36.63
19	8.82	19.29	34.51
20	9.24	18.79	
20			
20			
20			
Total	183.41	381.81	661.06
Mean	9.17	19.09	34.79
SDI	0.46	0.84	1.23
CV	0.05	0.04	0.04

CLIA acceptance performance for FSH is 25%.

Acceptable limit performance for RIQAS is 11,4%.