

FT3			
	Quality Control		
	Level 1	Level 2	Level 3
1	3.75	7.67	12.68
2	3.24	7.57	12.33
3	2.94	7.18	11.85
4	2.98	7.18	12.06
5	2.98	7.27	11.87
6	3.13	7.53	12.06
7	3.17	7.76	12.93
8	3.3	7.9	13.17
9	3.17	8.04	12.88
10	3.26	8.03	13.17
11	3.31	7.78	13.72
12	3.61	0.08	13.04
13	3.44	8.05	13.67
14	3.51	8.11	13.14
15	3.34	7.96	13.48
16	3.3	7.83	13.05
17	3.3	7.7	12.71
18	3.39	8.4	12.65
19	3.35	7.9	13.71
20	3.47	7.98	13.4
21	3.34	8.31	13.78
22	3.3	7.65	13.88
23	3.3	7.93	13.56
Total	75.88	171.81	298.79
Mean	3.30	7.47	12.99
SDI	0.19	1.64	0.63
CV	0.06	0.22	0.05

CLIA acceptance performance for %. FT3 is 25%
Acceptable limit performance for RIQAS is 13,6%.