

FT4			
	Quality Control		
	Level 1	Level 2	Level 3
1	12.1	30.73	44.54
2	12.64	32.7	47.67
3	12	30.02	44.56
4	11.31	30.54	44.64
5	11.23	29.33	44.96
6	11.17	29.7	44.23
7	11.51	30.37	43.3
8	11.7	31.08	44.31
9	11.52	29.73	43.76
10	11.72	29.85	45.04
11	12.81	33.78	50.97
12	12.53	33.76	49.68
13	12.33	34.02	51.41
14	12.74	35.22	51.16
15	12.49	35.5	54.66
16	13.06	34.99	52.11
17	12.99	35.38	54.74
18	12.8	34.24	52.06
19	12.76	34.02	49.88
20	12.68	33.87	50.64
21	13.01	33.65	50.85
22	12.48	33.32	50.81
23	11.16	30.32	46.65
Total	280.74	746.12	1112.63
Mean	12.21	32.44	48.38
SDI	0.65	2.16	3.71
CV	0.05	0.07	0.08

CLIA acceptance performance for FT4 is 25%.

Acceptable limit performance for RIQAS is 12,2%.