

Ferritin			
	Quality Control		
	Level 1	Level 2	Level 3
1	20.4	134	291.6
2	19.3	121.6	263.9
3	22.6	131.6	298.7
4	22.6	136.2	303.4
5	21.5	128.5	290
6	21.4	126.8	293
7	22.3	133	316.1
8	22.7	137.1	310.6
9	23.6	141.4	337
10	21.3	131.7	305.1
11	23.4	137.4	332.3
12	22.8	126.3	291.6
13	21	129.3	279.4
14	21.8	128.4	290.1
15	20.3	124.9	289.4
16	21.7	137.2	284.4
17	22	130.5	290.8
18	21.6	127	276.6
19	21.5	131.5	293.2
20	20.5	115.1	272.4
21	20.7	116	285.7
22	20.3	122.3	264.1
23			
Total	475.3	2847.8	6459.4
Mean	21.60	129.45	293.61
SDI	1.10	6.77	18.68
CV	0.05	0.05	0.06

CLIA acceptance performance for Ferritin is 25%.
Acceptable limit performance for RIQAS is 13,8%.