

LH			
	Quality Control		
	Level 1	Level 2	Level 3
1	4.85	21.8	63.21
2	4.31	19.84	62.81
3	4.39	20.31	64.9
4	4.08	20.06	63.41
5	4.53	20.65	63.04
6	4.34	20.74	62.26
7	4.55	20.46	63.92
8	4.35	19.37	63.58
9	5	20.7	63.81
10	4.74	21.5	61.9
11	4.9	21.95	65.98
12	4.71	21.66	64.28
13	4.72	21.28	61.9
14	4.51	20.89	65.98
15	4.64	21.68	64.28
16	4.61	21.78	66.09
17	4.41	20.56	
18	4.72	19.13	
19	4.5	21.19	
20	4.72	21.83	
21	4.51	20.89	
22	4.64	21.68	
23	4.4	21.78	
Total	105.13	2507.27	1021.35
Mean	4.57	109.01	63.83
SDI	0.21	422.25	1.37
CV	0.05	3.87	0.02

CLIA acceptance performance for LH is 25%.
Acceptable limit performance for RIQAS is 12%.