

<b>PSA</b>			
	<b>Quality Control</b>		
	<b>Level 1</b>	<b>Level 2</b>	<b>Level 3</b>
1	0.4	4.86	28.39
2	0.38	4.24	26.74
3	0.37	4.22	25.3
4	0.38	3.94	25.81
5	0.36	3.99	25.44
6	0.36	4.03	24.92
7	0.38	3.96	25.84
8	0.39	4.16	26.61
9	0.36	4.32	25.66
10	0.38	4.11	26.18
11	0.38	4.07	28.06
12	0.36	4.26	26.15
13	0.38	4.19	26.6
14	0.38	4.38	27.02
15	0.4	4.37	25.91
16	0.39	4.49	27.69
17	0.39	4.28	26.28
18	0.37	4.29	25.42
19	0.36	4.11	26.09
20	0.36	4.06	26.82
21	0.37	4.07	27.34
22			
23			
<b>Total</b>	<b>7.9</b>	<b>88.4</b>	<b>554.27</b>
<b>Mean</b>	<b>0.38</b>	<b>4.21</b>	<b>26.39</b>
<b>SDI</b>	<b>0.01</b>	<b>0.21</b>	<b>0.92</b>
<b>CV</b>	<b>0.04</b>	<b>0.05</b>	<b>0.03</b>

CLIA acceptance performance for PSA is 25%.

Acceptable limit performance for RIQAS is 12,3%.