

| <b>Progesterone</b> |                        |                |                |
|---------------------|------------------------|----------------|----------------|
|                     | <b>Quality Control</b> |                |                |
|                     | <b>Level 1</b>         | <b>Level 2</b> | <b>Level 3</b> |
| 1                   | 6.18                   | 47.2           | 109.35         |
| 2                   | 5.13                   | 42.35          | 84.49          |
| 3                   | 3.61                   | 31.89          | 91.58          |
| 4                   | 4.15                   | 31.79          | 97.84          |
| 5                   | 5.04                   | 37.98          | 104.33         |
| 6                   | 4.92                   | 38.39          | 104.79         |
| 7                   | 5.29                   | 40.13          | 106.72         |
| 8                   | 6.22                   | 38.47          | 100.71         |
| 9                   | 5.25                   | 40.1           | 89.41          |
| 10                  | 4.25                   | 34.66          | 96.31          |
| 11                  | 4.65                   | 34.27          | 94.79          |
| 12                  | 4.61                   | 34.19          | 98.28          |
| 13                  | 5.57                   | 37.21          | 95.38          |
| 14                  | 4.44                   | 38.29          | 92.6           |
| 15                  | 4.84                   | 36.88          | 98.86          |
| 16                  | 5.49                   | 36.86          | 107.36         |
| 17                  | 6.25                   | 38.29          | 109.92         |
| 18                  | 4.59                   | 36.88          | 98.34          |
| 19                  | 5.01                   | 36.86          |                |
| 20                  |                        | 36.98          |                |
| 21                  |                        |                |                |
| 22                  |                        |                |                |
| 23                  |                        |                |                |
| <b>Total</b>        | <b>95.49</b>           | <b>749.67</b>  | <b>1781.06</b> |
| <b>Mean</b>         | <b>5.03</b>            | <b>37.48</b>   | <b>98.95</b>   |
| <b>SDI</b>          | <b>0.71</b>            | <b>3.48</b>    | <b>7.11</b>    |
| <b>CV</b>           | <b>0.14</b>            | <b>0.09</b>    | <b>0.07</b>    |

CLIA acceptance performance for Prog is 25%.

Acceptable limit performance for RIQAS is 16,1%.