

Prolactin			
	Quality Control		
	Level 1	Level 2	Level 3
1	8.53	17.05	43.36
2	7.87	16.05	44.12
3	8.69	16.94	43.34
4	8.46	16.67	43.41
5	8.87	17.38	43.26
6	8.75	16.6	44.21
7	8.59	17.63	44.32
8	8.18	16.85	43.8
9	8.6	16.87	
10	8.39	17.3	
11	8.31	16.69	
12		17.05	
13		17.19	
14		17.42	
15		16.72	
16		16.76	
17			
18			
19			
20			
21			
22			
23			
Total	93.24	271.17	349.82
Mean	8.48	16.95	43.73
SDI	0.28	0.39	0.44
CV	0.03	0.02	0.01

CLIA acceptance performance for Prolactin is 25%.

Acceptable limit performance for RIQAS is 17,3%.