

<b>TSH</b>			
	<b>Quality Control</b>		
	<b>Level 1</b>	<b>Level 2</b>	<b>Level 3</b>
1	0.789	5.913	32.62
2	0.775	5.62	28.17
3	0.819	6.167	35.25
4	0.854	6.732	38.45
5	0.846	6.413	36.86
6	0.78	5.455	30.78
7	0.724	5.433	28.46
8	0.726	5.132	29.12
9	0.708	4.992	29.18
10	0.747	5.328	29.44
11	0.726	5.06	28.7
12	0.717	5.202	29.97
13	0.763	5.622	30.41
14	0.708	5.251	28.45
15	0.688	5.156	29.02
16	0.702	5.329	29.06
17	0.743	5.542	29.58
18	0.688	5.458	30.46
19	0.733	5.294	28.27
20	0.743	5.542	29.588
21	0.688	5.458	30.463
22	0.733	5.294	28.27
23	0.693	5.211	28.997
<b>Total</b>	<b>17.093</b>	<b>126.604</b>	<b>699.568</b>
<b>Mean</b>	<b>0.74</b>	<b>5.50</b>	<b>30.42</b>
<b>SDI</b>	<b>0.05</b>	<b>0.43</b>	<b>2.79</b>
<b>CV</b>	<b>0.06</b>	<b>0.08</b>	<b>0.09</b>

CLIA acceptance performance for TSH is 25%.

Acceptable limit performance for RIQAS is 11,3%.