

Testosterone			
	Quality Control		
	Level 1	Level 2	Level 3
1	3.26	12.68	23
2	3.41	14.32	30.57
3	3.35	12.9	27.38
4	3.35	12.89	26.16
5	3.48	13.07	26.65
6	3.8	12.92	26.96
7	3.66	12.97	30.52
8	3.89	14.34	30.54
9	3.9	14.39	30.26
10	3.91	14.39	30.4
11		14.57	31.13
12		14.58	30.64
13		14.48	30.6
14		14.48	31.16
15		14.16	31.54
16		14.9	31.32
17		14.66	
18			
19			
20			
21			
22			
23			
Total	36.01	236.7	468.83
Mean	3.60	13.92	29.30
SDI	0.26	0.79	2.48
CV	0.07	0.06	0.08

CLIA acceptance performance for Testo is 25%.

Acceptable limit performance for RIQAS is 15,9%.