

Vitamin B12			
	Quality Control		
	Level 1	Level 2	Level 3
1	171	359	471
2	144	326	445
3	133	268	380
4	127	281	370
5	130	27	354
6	127	295	361
7	129	265	367
8	131	268	350
9	125	287	379
10	136	282	384
11	141	275	397
12	119	293	373
13	125	296	369
14	124	280	359
15	121	295	382
16	125	283	396
17	136	324	368
18	141	324	474
19	140	314	442
20			
21			
22			
23			
Total	2525	5342	7421
Mean	132.89	281.16	390.58
SDI	11.72	66.18	38.43
CV	0.09	0.24	0.10

CLIA acceptance performance for Vit B12 is 25%.

Acceptable limit performance for RIQAS is 17%.