

INTER-LABORATORY COMPARISON

Doc no	PRL- ILC Form	Revision 0.0	Effective Date: 1 Jan 2021
Written By: Gugu Ditshego	Reviewed by: Gugu Ditshego		Approved by: Jabulani Kubheka
Date: Dec 2020	Date: Dec 2020	Date: January 2021	



12th Road & Greer Road
Springs
1563
Tel: 0604707089
Email: info@proclinicalabs.co.za



Proteus
Laboratories
In pursuit of perfection

Number of Samples	2
Frequency	3-monthly
Test method (manual)	ABO & Rh

Date of running	Participant Result	Comparison Result	% Agreement	Comment
04/03/2021	A Negative	A Negative	100%	149560
04/03/2021	O Positive	O Positive	100%	149508
04/03/2021	B Positive	B Positive	100%	149538

The laboratories much reach a consensus for ILC to be acceptable. When there is no consensus the sample needs to be sent to a 3rd laboratory.

Corrective action No action required

Comment Ditshego 100% acceptable

Jabulani Kubheka